



Australian Government

**Assessment Requirements for
SHBBHRS005 Identify and control safety
risks for intense pulsed light and laser hair
reduction treatments**

Release: 1

Assessment Requirements for SHBBHRS005 Identify and control safety risks for intense pulsed light and laser hair reduction treatments

Modification History

Not applicable.

Performance Evidence

Evidence of the ability to complete tasks outlined in elements and performance criteria of this unit in the context of the job role, and:

- identify three safety hazards and complete three risk assessments in at least one of the following:
 - an intense pulsed light treatment environment
 - a laser treatment environment
- develop, implement and maintain safety practices and risk control measures to respond to safety hazards associated with use of at least one of the following:
 - intense pulsed light
 - laser
- for the above, record the following:
 - workplace documentation and checks for the maintenance of equipment
 - workplace health and safety reporting related to intense pulsed light or laser hair reduction treatments
- during risk assessments, follow workplace policies and procedures, non-ionising radiation safety protection plan and relevant state or territory laws, codes and guidelines related to the use of intense pulsed light and laser for hair reduction.

Knowledge Evidence

Demonstrated knowledge required to complete the tasks outlined in elements and performance criteria of this unit:

- key aspects of current legislation, regulation and industry standards and guidelines relevant to intense pulsed light and laser hair reduction treatments:
- laser classification and hazard analysis
- maintenance and auditing of facilities
- equipment inspection and maintenance protocols
- investigation and management of intense pulsed light and laser incidents
- personal protective equipment
- potential for fire and explosion and protection against flammability hazards
- management of airborne contaminants

- electrical safety laser controlled treatment areas:
 - designation
 - warning signs
 - entry controls
 - control of access to the laser
- safety of intense pulsed light and laser products
- equipment classification
- safe use of intense pulsed light and laser equipment
- intense pulsed light and laser hazards:
 - beam hazards
 - non-beam hazards
- risk and hazard management including risk assessment
- hierarchy of hazard control:
 - engineering controls
 - administration controls
 - personal protective equipment
 - quality assurance testing and preventive maintenance
 - safe work practices
- workplace work health and safety policies:
 - procedures for:
 - identifying hazards
 - assessing risk
 - controlling risk
 - non-ionising radiation safety protection plan
 - ergonomics
- types of safety hazards related to intense pulsed light and laser hair reduction and associated control measures:
 - ocular and skin
 - equipment specific
 - reflected beams
 - fire, explosive, electrical and environmental
- control measures applied to minimise health and safety risks during intense pulsed and laser hair reduction treatments:
 - pre-treatment patch testing
 - personal protective equipment:
 - eye protection
 - masks
 - gowns
 - gloves
 - ventilation systems

- filtering and exhaust systems
- window coverings
- draping of treatment environment
- equipment preventive maintenance:
 - documentation, including:
 - equipment safety standards
 - equipment service history
 - schedule for recommended testing and maintenance
 - regular testing
- biophysics of laser and intense pulsed light technologies as outlined in *AS/NZS 4173:2018, Safe use of lasers and intense light sources in health care* or its replacement
- how light energy interacts with skin and hair
- physics of light related to intense pulsed light and laser treatments:
 - energy forms
 - electromagnetic spectrum
 - optical region of the electromagnetic spectrum
 - wavelengths
 - characteristics of a wave
- operational characteristics of lasers:
 - programmable systems
 - basic set-up (manual choice)
 - generation of laser beam and propagation of light
 - characteristics of laser beams
 - properties of different types of lasers
 - optical pathways
 - delivery systems and applicators used for hair reduction
- operational characteristics of intense pulsed light apparatus:
 - programmable systems
 - basic set-up (manual choice)
 - powerful systems
 - significantly lower power systems
 - common differences in imported systems
 - chilled sapphire or similar optical substance head
 - non-chilled sapphire head
 - non-laser light source
 - characteristics of flashlamp
 - capacitors - free discharge and partial discharge
 - different filters
 - properties of intense pulsed light equipment
- laser controls:

- emergency control
- delivery systems
- ionising and non-ionising radiation
- parameters that effect the delivery of light:
 - spectrum of delivered wavelengths as determined by cut-off filters
 - number of delivered pulses, including single and multiple pulsed shots
 - pulse duration in milliseconds
 - delay between pulses in milliseconds
 - delivered fluence
 - laser hazards.

Assessment Conditions

Skills must be demonstrated in:

- an industry workplace, or
- a simulated industry environment.

Assessment must ensure access to:

- documentation related to intense pulsed light and laser treatments:
 - workplace policies and procedures
 - workplace reporting documentation and templates
 - current standards and guidelines:
 - AS/NZS 4173:2018 Safe use of lasers and intense light sources in health care
 - AS/NZS IEC 60825.1:2014 Safety of laser products – Part 1: Equipment classifications and requirements
 - AS/NZS 1336:2014 Eye and face protection - Guidelines
 - key aspects of relevant local, federal, state or territory, legislation and regulations relating to use of intense pulsed light and laser for hair reduction
 - manufacturer instructions and safety data sheets
 - non-ionising radiation safety protection plan
- a clinical treatment area:
 - as per relevant state or territory legislation either, or both, of the following:
 - a radiation warning sign
 - an illuminated light above door
 - non-flammable screens fitted inside all windows to protect a person outside window from non-ionising radiation levels greater than maximum permissible exposure from radiation
 - ventilation designed to ensure that airborne hazards are not passed downstream in air handling or exhaust system
 - fire extinguishing equipment
 - sufficient clinical lighting

- eye protection equipment for practitioners and clients compliant with AS/NZS 1336:2014 Eye and face protection – Guidelines
- disposable masks
- equipment which, when energised, can emit an amount of non-ionising radiation higher than accessible limit for a Class 3B laser for relevant period stated in, and measured in accordance with, laser standards AS2211. This must include a Class 4 laser or equivalent in the case of IPL.
- one or more of the following:
 - cosmetic laser equipment that:
 - can treat Fitzpatrick skin types one to six
 - is registered for purpose on the Australian Register of Therapeutic Goods
 - intense pulsed light equipment that is registered for purpose on the Australian Register of Therapeutic Goods and has one or more of the following characteristics:
 - programmable
 - manual
 - multiple pulsed shots
 - single pulsed shots
 - appropriate cooling delivery systems
 - one or minimal choice of filters
 - multiple filters
- activities that require the individual to work within commercially realistic timing and productivity.

Assessors must satisfy the Standards for Registered Training Organisations' requirements for assessors, and:

- have at least three years' experience in operating laser equipment; **OR**
- have at least three years' experience in operating intense pulsed light equipment.

Links

Companion Volume Implementation Guide -

<https://vetnet.gov.au/Pages/TrainingDocs.aspx?q=255d312b-db07-48f2-b6d6-1b0b06c42898>